CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD







0654



Unit 5 Octavian Way Team Valley Trading Estate Gateshead NE11 0HZ

Tel: +44 (0)191 4875951

Page 1 of 3							
Approved Signate	ory						
Electronically A	authorised Document						
P K CLARK R J WADE	☐ J FRYER ☐ M FOY						
M A FROST M S PARDOE							

CUSTOMER

DJB LABCARE LTD UNIT 12 HOWARD WAY CROMWELL BUSINESS CENTRE NEWPORT PAGNELL BUCKINGHAMSHIRE MK16 9QS UNITED KINGDOM MANUFACTURER STANDARD

DESCRIPTION TACHOMETER OPTICAL

MODEL AT-6

SERIAL No. 200552064

IDENT No. UNKNOWN

DATE RECEIVED 4 OCTOBER 2024

DATE OF CALIBRATION 9 OCTOBER 2024

ORDER No. NS23936

INSTRUMENT CONDITION

Adjustments Made No Repairs Made No

ENVIRONMENT

The instrument was placed in the Laboratory environment for a minimum period of 4 hours prior to calibration.

The ambient conditions were: 18 to 23°C and 45 %rh ± 15 %rh.

STABILITY

The results contained in this Certificate refer to the measurements made at the time of test and not to the instrument's ability to maintain calibration.

PROCEDURE

Measurements were performed in accordance with the in house Laboratory procedure No.1053

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k = 2, providing a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD



CERTIFICATE NUMBER

1169117

Page 2 of 3

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654

Calibration Equipm	Test Equipment			
Cert Number	Ident Number	Model	Serial Number	Calibration Due
1168575IH	143	33220A	MY44025316	3 Oct 2025
1149140IH	475	A002206.G1	13572091	3 Sep 2026

Notes:

Results relate only to the items calibrated.

CERTIFICATE OF CALIBRATION

ISSUED BY: CALIBRATION MAINTENANCE & REPAIR LTD



UKAS ACCREDITED CALIBRATION LABORATORY No. 0654

RESULT SHEET 1053-NC-UKAS: STANDARD AT-6 OPTICAL TACHOMETER

AS FOUND

BATTERY REPLACED: NO

The Revolution Per Minute (RPM) was simulated using a precision Frequency Generator with an attached optical light source.

RPM

Range	Simulated	Indicated	Units	Indicated Deviation	Measurement Uncertainty ±
2 to	600.000	600.0	RPM	0.000	0.064
99999RPM	1000.00	1000	RPM	0.00	0.58
	2500.00	2500	RPM	0.00	0.58
	5000.00	5000	RPM	0.00	0.58
	7500.00	7500	RPM	0.00	0.58
	9999.98	10000	RPM	0.02	0.58
	14999.65	15000	RPM	0.35	0.58
	19999.43	20000	RPM	0.57	0.58
	24999.13	25000	RPM	0.87	0.58
	29998.84	30000	RPM	1.16	0.58

Manufacturer stated Accuracy: ±(0.05%+1 digit)

COMMENTS:

TEST ENGINEER: M Foy DATE: 09th October 2024

Issue No 1 Page 3 of 3