

Certificate of Calibration

Certificate number: **C1505-289**Equipment: TIM 901R Stopwatch
Serial number: 001Date: 01/05/2015
Calibration frequency: Every 12 months

Calibrated by: Martin Paragreen

Results:

Set Value	Measured Value
05:00.00	04:59.96
10:00.00	09:59.84
60:00.00	59:59.89

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: P Smith

