

CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

DATE OF ISSUE: 3 March 2016

CERTIFICATE No: 387626




Lambda
CALIBRATION LTD

Units 11 - 13
Chorley Central Business Park
Stump Lane, Chorley
Lancashire PR6 0BL
Tel: 0845 2411533
Fax: 0845 2411544

Page 1 of 2

APPROVED SIGNATAORY


A Kelly D Pilkington
D Whalley C Reed R Armitage

Customer: DJB Labcare Ltd
Address: 20 Howard Way, Interchange Park,
Milton Keynes
MK16 9QS

Item Number: 12078446 (4046)
Description: Digital Multimeter
Model/Range: AT-6
Manufacturer: Standard
Date of Cal: 3 Mar 2016
Calibrated by: Jacob Rimmer
Procedure Name: Standard, Digital Tachometer, AT-6
Rev/Basis: 01:E-1000
Temp/Humidity: 20.0°C ±2°C <80%rh

The Results on the following pages are: As Found

All Measurements are Traceable to National Standards.

Note 1: The unit under test was calibrated using a multifunction calibrator.
Note 2: Where the reported value lies within the specified tolerances then this will be indicated by the word "PASS", if outside then by the word "FAIL".
Note 3: Values quoted in the "UUT Indicated Value" column are not necessarily quoted to the same resolution as the actual displayed value on the UUT.
Note 4: Any supplied test leads have been checked as part of the Visual/Operational test but have not been used during calibration.

Engineers' Notes:

Standard(s) Used: LMMC-02 / LMMC-04 / LMMC-10 // LMMC-14

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

ISSUED BY: LAMBDA CALIBRATION LTD

UKAS ACCREDITED CALIBRATION LABORATORY No: 0495

CERTIFICATE No:

387626

Page 2 of 2

Parameter	UUT Range	Applied Value	UUT Reading	Acceptance Limits Low	Acceptance Limits High	Pass/Fail
-----------	-----------	---------------	-------------	-----------------------	------------------------	-----------

Visual/Operational Test

Result of Operator Evaluation

PASS

Various frequencies were applied from a multi-function calibrator to the tachometer optical calibration box, LSTR-06, the output of which was measured with the UUT. The results were as follows:

RPM	Applied Value	UUT Reading	Acceptance Limits Low	Acceptance Limits High	Pass/Fail
600.0rpm	600.0	599.6	600.4		PASS
1000rpm	1000	999	1002		PASS
2500rpm	2500	2498	2502		PASS
5000rpm	5000	4997	5004		PASS
7500rpm	7500	7495	7505		PASS
10000rpm	10000	9994	10006		PASS
14000rpm	14000	13992	14008		PASS
18000rpm	18000	17990	18010		PASS
20000rpm	19999	19989	20011		PASS
25000rpm	24999	24987	25014		PASS
30000rpm	30001	29984	30016		PASS
40000rpm	39998	39979	40021		PASS
50000rpm	50001	49974	50026		PASS
60000rpm	60003	59969	60031		PASS
70000rpm	70000	69964	70036		PASS
80000rpm	80002	79959	80041		PASS
90000rpm	90005	89954	90046		PASS

End of Calibration Data

Estimated Uncertainty of Measurement:

Frequency

0.2Hz to 250MHz: +/- (0.29ppm + 2 LSD)