



Certificate of Calibration

Certificate number: 190316

Equipment: Fastime 3 Stopwatch
Serial number: 019

Date: 11/03/2016
Calibration frequency: Every 12 months

Calibrated by: Martin Paragreen

Results:

Set Value	Measured Value
05:00.00	05:00.00
10:00.00	09:59.96
1:00:00.00	59:59.92

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: P Smith