



# Certificate of Calibration

Certificate number: 180818

Equipment: Marble Stopwatch  
Serial number: 018

Date: 22/08/18  
Calibration frequency: Every 12 months

Calibrated by: Caitriona Hurd

## Results:

Set Value	Measured Value
05:00.00	05:00.04
10:00.00	10:00.04
60:00.00	59:59.92

Method of test:  
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam