



Certificate of Calibration

Certificate number: 12082019

Equipment: Fastime 3 Stopwatch
Serial number: 019

Date: 12/08/2019
Calibration frequency: Every 12 months

Calibrated by: Jude Huckle

Results:

Set Value	Measured Value
05:00.00	04:59.86
10:00.00	10:00.00
60:00.00	60:00.10

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam