



Certificate of Calibration

Certificate number: 20/111019

Equipment: Fastime 3 Stopwatch
Serial number: 020

Date: 11/10/19
Calibration frequency: Every 12 months

Calibrated by: Jude Huckle

Results:

Set Value	Measured Value
05:00.00	05:00.12
10:00.00	10:00.09
60:00.00	60:00.06

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam

