

CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

DATE OF ISSUE: 18 October 2021 CERTIFICATE NUMBER: **1118000**



0654



Unit 5 Octavian Way
Team Valley Trading Estate
Gateshead
NE11 0HZ

Tel: +44 (0)191 4875951

Page 1 of 4

Approved Signatory

Electronically Authorised Document

- | | |
|--|----------------------------------|
| <input type="checkbox"/> P K CLARK | <input type="checkbox"/> J FRYER |
| <input type="checkbox"/> R J WADE | <input type="checkbox"/> M FOY |
| <input type="checkbox"/> M A FROST | |
| <input checked="" type="checkbox"/> M S PARDOE | |

CUSTOMER

DJB LABCARE LTD
UNIT 12 HOWARD WAY
CROMWELL BUSINESS CENTRE
NEWPORT PAGNELL
BUCKINGHAMSHIRE
MK16 9QS
UNITED KINGDOM

MANUFACTURER

DESCRIPTION

MODEL

SERIAL No.

IDENT No.

DATE RECEIVED

DATE OF CALIBRATION

ORDER No.

STANDARD

TACHOMETER OPTICAL

AT-6

12078389

UNKNOWN

15 OCTOBER 2021

18 OCTOBER 2021

NS23416

INSTRUMENT CONDITION

Adjustments Made No
Repairs Made No

ENVIRONMENT

The instrument was placed in the Laboratory environment for a minimum period of 4 hours prior to calibration.

The ambient conditions were: 22°C ± 3°C and 45% RH ± 15% RH.

STABILITY

The results contained in this Certificate refer to the measurements made at the time of test and not to the instrument's ability to maintain calibration.

PROCEDURE

Measurements were performed in accordance with the in house Laboratory procedure No.1053

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**



CERTIFICATE NUMBER

1118000

Page 2 of 4

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654

INSTRUMENTS USED **EQUIPMENT**

Agilent 33220A
Trimble A002206.G1

SERIAL No
MY44025316
13431078

CERTIFICATE No	CAL DUE
M3198	30 Sep 2022
M2988	08 Sep 2023

Notes:

Results relate only to the items calibrated.

Measurement Uncertainties

These are our best measurement capabilities with the listed test equipment, any uncertainties shown on the result sheet take in to account the resolution of the instrument being calibrated.

Parameter	Range	Uncertainty	Parameter	Range	Uncertainty
Revolutions Per Minute	6RPM to 600RPM	0.064RPM			
	600RPM to 6000RPM	0.065RPM			
	6000RPM to 120000RPM	0.066RPM			

CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654



CERTIFICATE NUMBER

1118000

Page 3 of 4

Parameter

Range

Uncertainty

Parameter

Range

Uncertainty

CERTIFICATE OF CALIBRATION

ISSUED BY: **CALIBRATION MAINTENANCE & REPAIR LTD**

UKAS ACCREDITED CALIBRATION LABORATORY No. 0654



CERTIFICATE NUMBER

1118000

RESULT SHEET 1053-NC-UKAS : STANDARD AT-6 OPTICAL TACHOMETER

AS FOUND

BATTERY REPLACED: NO

The Revolution Per Minute (RPM) was simulated using a precision Frequency Generator with an attached optical light source.

RPM

Range	Simulated	Indicated	Units	Indicated Deviation	Measurement Uncertainty \pm
2 to 99999RPM	600.000	600.0	RPM	0.000	0.064
	1000.00	1000	RPM	0.00	0.58
	2500.00	2500	RPM	0.00	0.58
	5000.00	5000	RPM	0.00	0.58
	7500.00	7500	RPM	0.00	0.58
	10000.00	10000	RPM	0.00	0.58
	14999.46	15000	RPM	0.54	0.58
	19999.14	20000	RPM	0.86	0.58
	24998.76	25000	RPM	1.24	0.58
	29998.44	30000	RPM	1.56	0.58

Manufacturer stated Accuracy: $\pm(0.05\%+1 \text{ digit})$

COMMENTS:

TEST ENGINEER: J Fryer

DATE: 18th October 2021