



12 Howard Way
Cromwell Business Centre
Newport Pagnell
Bucks MK16 9QS

Tel: 01908 612598
Fax: 01908 217974

E-mail: service@djblabcare.co.uk
Website: www.djblabcare.co.uk

Certificate of Calibration

Certificate number: 20-300922

Equipment: Fastime 3 Stopwatch
Serial number: 020

Date: 30/09/2022
Calibration frequency: Every 12 months

Calibrated by: Jude Huckle

Results:

Set Value	Measured Value
05:00.00	05:00.04
10:00.00	10:00.03
60:00.00	60:00.04

Method of test:

All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam