



## Certificate of Calibration

Certificate number: 15/220223

Equipment: Fastime 3 Stopwatch  
Serial number: 015

Date: 22/02/2023  
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

### Results:

Set Value	Measured Value
05:00.00	04:59.90
10:00.00	10:00.09
60:00.00	59:59.80

Method of test:  
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam

