



# CERTIFICATE OF CALIBRATION

Issued By

**ABSOLUTE CALIBRATION LIMITED**



0078

DATE OF ISSUE 29 May 2024

CERTIFICATE NO. 0541040



## Absolute Calibration Limited

14 Murrills Estate, Portchester  
Hampshire, England PO16 9RD

Telephone: 023 9232 1712

[www.absolutecal.co.uk](http://www.absolutecal.co.uk)

Page 1 of 2

Approved Signatory

D Kingswell

G Mills

S Patabendi

A Watson

Manufacturer: TTI  
Type Number: PFM1300  
Description: High Resolution Frequency Counter  
Serial Number: 230096  
Customer Reference: 513815-2  
Customer Code: PUL001  
Customer: Pullman Instruments (UK) Limited  
ESG House  
Chatsworth Road  
Harrogate  
North Yorkshire

On Behalf Of: DJB Labcare Limited

Order Number: 258108

Instrument Receipt Date: 15 May 2024

Laboratory Temperature: 20.0 °C ± 3.0 °C

Laboratory Humidity: 50 %rh ± 25 %rh

Unit Stabilisation Time: Twenty-Four Hours

Calibration Procedure: CP2006

Calibration Engineer: J. Perkis

Calibration Date: 28 May 2024

This report contains: Recorded results with no adjustments

Pre and post adjustment results

Post repair results

Results recorded at Customer site

The following calibration results relate to the items defined above or uniquely identified in the following pages.

# CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

0541040

UKAS Accredited Calibration Laboratory No. 0078

Page 2 of 2 Pages

## PARAMETER TESTED

### 'X' Tal Accuracy

<u>Applied</u> <u>Frequency</u>	<u>Uncertainty ± of</u> <u>Applied Value</u>	<u>Deviation from</u> <u>Applied Frequency</u>
10.000 000 000 MHz	2 in 10 <sup>9</sup>	< 2 in 10 <sup>6</sup>

The internal 'X' Tal oscillator was checked by applying a standard 10 MHz signal to the input of the counter and evaluating the resultant reading.

### Timebase Accuracy

<u>Applied</u> <u>Input</u>	<u>Uncertainty ± of</u> <u>Applied Input</u>	<u>Gate</u> <u>Time</u>	<u>PFM3000</u> <u>Display</u>
10.000 00 MHz	2 in 10 <sup>9</sup>	0.3 S	10.0000 MHz
10.000 000		1	10.00001
10.000 000 0		10	10.000014

### Frequency Response Channel A

<u>PFM3000 Setting</u>	<u>Applied</u> <u>Input</u>	<u>Uncertainty ± of</u> <u>Applied Input</u>	<u>PFM3000 Display</u> <u>Display</u>
1 s	10.000 000 Hz	2 in 10 <sup>9</sup>	10.00001 Hz
	50.000 000		50.00006
	100.000 000		100.0001
	300.000 000		300.0003
	500.000 000		500.0006
	700.000 000		700.0009
0.1 s	1.000 000 MHz	2 in 10 <sup>9</sup>	1.00000 MHz
	5.000 00		5.00000
	10.000 00		10.0000
	25.000 00		25.0002

### Frequency Response Channel B

<u>PFM3000 Setting</u>	<u>Applied</u> <u>Input</u>	<u>Uncertainty ± of</u> <u>Applied Input</u>	<u>PFM3000 Display</u> <u>Display</u>
0.1 s 20 MHz	20.000 0 MHz	2 in 10 <sup>9</sup>	20.0000 MHz
0.1 s 100 MHz	100.000 0		100.000
0.1 s 1.3 GHz	1 300.000	2 in 10 <sup>9</sup>	1300.001 MHz

An additional uncertainty of 1 lsd for the resolution of the display should be calculated using summation in quadrature.

The uncertainties reported after measured values only, with no account being taken of the instrument's ability to maintain its calibration.

--- End ---