

# Certificate of Calibration

Certificate number: **C1505-283**Equipment: TIM 901S Stopwatch  
Serial number: 004Date: 01/05/2015  
Calibration frequency: Every 12 months

Calibrate by: Martin Paragreen

Results:

Set Value	Measured Value
05:00.00	05:00.02
10:00.00	10:00.13
60:00.00	60:00.08

Method of test:  
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: P Smith

