



20 Howard Way
Cromwell Business Centre
Interchange Park
Newport Pagnell
Bucks MK16 9QS

Tel: 01908 612598
Fax: 01908 217974

E-mail: service@djblabcare.co.uk
Website: www.djblabcare.co.uk

Certificate of Calibration

Certificate number: DJB203-5

Equipment: G100 Gas Analyser
Serial number: INO4410

Analysers are calibrated on the morning of any pending CO2 calibrations in accordance with DJB Labcare procedures

Method

All calibrations are carried out in a temperature stable laboratory using reference gases providing traceability to UKAS ISO17025 standards and in accordance with the manufacturers procedures

Certified traceable sample gas from EfecTech, Laboratory Number 0590 Co2 5% - Cylinder number D690633. Certificate Number: 18/0796/01

Date of calibration 12th June 2018.

Approved by: Colin Bradnam

A handwritten signature in purple ink, appearing to read 'Colin Bradnam', written over a faint dotted grid.