



Certificate of Calibration

Certificate number: 311021

Equipment: Fastime 3 Stopwatch
Serial number: 021

Date: 31/10/18
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

Results:

Set Value	Measured Value
05:00.00	04:59.96
10:00.00	10:00.01
60:00.00	59:59.93

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam