



Certificate of Calibration

Certificate number: 16/5219

Equipment: Fastime 3 Stopwatch
Serial number: 016

Date: 05/02/19
Calibration frequency: Every 12 months

Calibrated by: Jude Huckle

Results:

Set Value	Measured Value
05:00.00	05:00.09
10:00.00	10:00.16
60:00.00	60:00.12

Method of test:
All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam