

CERTIFICATE OF CALIBRATION

ISSUED BY: **M K I S** CALIBRATION COMPANY

DATE OF ISSUE: 01 October 2010

CERTIFICATE NUMBER: 2376



0236

STANDARDS LABORATORY

M K I S



Calibration Company

10 Potters Lane
Kiln Farm

Milton Keynes
MK11 3HE

Tel: 01908 568250
Fax: 01908 564661

Page 1 of 2 pages

Approved
Signatories

R Younger
C Kemp

Signature

Equipment Description:	Tachometer
Manufacturer:	Standard
Type:	ST 6236B
Serial Number:	08013379
Order Number:	17035
Customer:	D J B Labcare
Location:	Newport Pagnell
Date Received:	29 September 2010
Date Calibrated:	30 September 2010

The instrument was kept in the laboratory environment for 24 hours, to allow the instrument to stabilise, prior to the tests being carried out.

The ambient temperature and relative humidity throughout the test was $20^{\circ}\text{C} \pm 2^{\circ}\text{C}$ and $50\% \pm 20\%$ respectively.

The uncertainties reported refer to the applied values only with no account being taken of the instruments ability to maintain its calibration.

Remarks: No adjustments were made.

CERTIFICATE OF CALIBRATION

Certificate Number: 2376

UKAS Accredited Calibration Laboratory No. 0236

PAGE 2 OF 2 PAGES

Applied Value	Equivalent Value	Indicated Value
100.18 ms	598.92 RPM	598.8 RPM
60.020 ms	999.67 RPM	999.6 RPM
30.008 ms	1999.47 RPM	1999 RPM
20.002 ms	2999.7 RPM	2999 RPM
15.002 ms	3999.5 RPM	3999 RPM
12.006 ms	4997.5 RPM	4997 RPM
10.001 0 ms	5999.4 RPM	5999 RPM
7.500 08 ms	7999.9 RPM	8000 RPM
5.999 62 ms	10 000.6 RPM	10 001 RPM
3.000 42 ms	19 997.2 RPM	19 997 RPM
1.500 08 ms	39 997.9 RPM	39 998 RPM
1.000 07 ms	59 995.8 RPM	59 996 RPM

The measurement uncertainties were:

Time $\pm 0.01\% + 1 \text{ LSD}$

END

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.