



12 Howard Way  
Cromwell Business Centre  
Newport Pagnell  
Bucks MK16 9QS

Tel: 01908 612598  
Fax: 01908 217974

E-mail: [service@djablabcare.co.uk](mailto:service@djablabcare.co.uk)  
Website: [www.djablabcare.co.uk](http://www.djablabcare.co.uk)

## Certificate of Calibration

Certificate number: 10/210622

Equipment: Fastime 3 Stopwatch  
Serial number: 010

Date: 21/06/2022  
Calibration frequency: Every 12 months

Calibrated by: Colin Bradnam

### Results:

Set Value	Measured Value
05:00.00	05:00.03
10:00.00	10:00.09
60:00.00	59:59.96

### Method of test:

All calibrations are carried out using an in house calibration work instruction WI40

Expanded uncertainty is 0.48s

Approved by: C Bradnam